




# UNITED STATES PATENT AND TRADEMARK OFFICE

*JFL*




UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILED DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/803,801	03/18/2004	Sascha Albus	LZ-85	5607
7590 09/05/2007				
Friedrich Kueffner 317 Madison Avenue, Suite 910 New York, NY 10017				
			EXAMINER	
			JACKSON, BRANDON LEE	
			ART UNIT	PAPER NUMBER
			3772	
			MAIL DATE	DELIVERY MODE
			09/05/2007	PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

**Office Action Summary**



Application No.

10/803,801

Applicant(s)

ALBUS ET AL.

Examiner

Brandon Jackson

Art Unit

3772

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 02 July 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-12 is/are pending in the application.
- 4a) Of the above claim(s) 10-12 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-9 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

### **DETAILED ACTION**

This action is in response to arguments filed 7/2/2007. Currently, claims 1-12 are pending in the instant application, wherein claims 10-12 have been withdrawn from consideration.

#### ***Response to Arguments***

Applicant's arguments, see pages 4-6, filed 7/2/2007, with respect to the rejection(s) of claim(s) 1-4 under 102(e) have been fully considered and are persuasive. Therefore, the rejection has been withdrawn. However, upon further consideration, a new ground(s) of rejection is made in view of Johns (US Patent 4,513,739).

#### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-5 and 7-9 are rejected under 35 U.S.C. 102(b) as being anticipated by Johns (US Patent 4,513,739). Johns discloses a wound-dressing product (100) comprising a wound dressing element (102) having an adhesive surface (104), a cover arrangement (106, 108, 610) having at least two cover elements (106), wherein the two cover elements (4) are separated by a partition line (112) and each cover element (106) has a removal aid (108) connected to the cover element (106) and where the two

Art Unit: 3772

removal aids (108) overlap (fig. 2) the two cover elements (106). Adhesive strips (col. 3, lines 41-45) extend parallel to the partition line (112) under the removal strips, fastening the removal strips to the cover elements (106). The removal strips (fig. 2) are separated by the partition line (112). The removal aid (108) and cover elements (106) may be made of the same material or made integrally, therefore, it can inherently be made of different materials. The haptic properties of different materials would allow the user to distinguish the removal aid (108) from the cover element (106). The cover arrangement (106, 108, 610) encloses the wound dressing element (102) and comprises two reinforcement elements (610) on the lateral edges of the wound dressing element (102).

### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claim 6 is rejected under 35 U.S.C. 103(a) as being unpatentable over Johns (US Patent 4,513,739). Johns substantially discloses the claimed invention; see rejections to claims 1 and 5 above. Johns fails to disclose the removal aid is distinguishable by color from the cover element. However, at the time of the invention it would have been an obvious matter of design choice to a person of ordinary skill in the art to make the removal aid and cover elements of different colors because Applicant has not disclosed that the different colors provide an advantage, is used for a particular purpose, or solves a stated problem. Therefore, it would have been obvious to one of ordinary skill in the art at the time of the invention to modify the color of the removal aid and cover elements of Johns to obtain the claimed invention because such a modification would have been considered a mere design consideration which fails to patentably distinguish over the prior art of Johns.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brandon Jackson whose telephone number is (571)272-3414. The examiner can normally be reached on Monday - Friday 8-5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Patricia Bianco can be reached on (571)272-4940. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3772

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Brandon Jackson  
Examiner  
Art Unit 3772

BLJ

*Patricia Bianco*  
PATRICIA BIANCO  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 3700  
9/3/07

Organization

IC3/00 RANDOLPH

Bidg. Room

U. S. DEPARTMENT OF COMMERCE  
COMMISSIONER FOR PATENTS

P.O. BOX 1450

ALEXANDRIA, VA 22313-1450

IF UNDELIVERABLE RETURN IN TEN DAYS

OFFICIAL BUSINESS

AN EQUAL OPPORTUNITY EMPLOYER

